

**SIMPLE ONE-STEP CALIBRATION
OPTION
(FOR RESP SENSE CAPNOGRAPH)**

Zero Baseline Adjust Method

1. Turn power on
2. Press audible alarm on
3. Select adult or pediatric
4. Now press audible alarm off
5. Simultaneously press alarm off (for total of 15 seconds), And press on button (once only)
6. Screen should display the last instruction
7. After 15 seconds, screen should say "calibrating"
8. Entire process is complete in 15 minutes